

Certificate Of Analysis



Client:
Dialed Research

Laboratory:

Liquilabs s.r.o.
Ohradní 24B
14000 Michle
Czechia
www.liquilabs.cz

Accredited Testing Laboratory
ISO/IEC 17025:2017
ČIA Reg. No. 1437

Sample Identification

Sample Name	Bacteriostatic Water 10 mL	Batch Number	468125	Date Published	2026-08-07 11:30
--------------------	----------------------------	---------------------	--------	-----------------------	------------------

Results for Bacteriostatic Water (0.9% Benzyl Alcohol)

Analysis of Peptide Identity, Content and Purity	Result Unit	Reporting Uncertainty Limit
Benzyl Alcohol Assay USP Sterility Testing	0.91 %	[± 0.02]
Sterility (USP <71>) USP Sterility Testing	PASS	No growth
Endotoxin (USP <85>) USP Sterility Testing	< 0.125 EU/mL	[± 0.01]
pH USP Sterility Testing	5.42	[± 0.05]

Responsibles

A handwritten signature in black ink, appearing to read "Galbavy", is positioned above the printed name and title.

Mr. Ján Galbavý
Founder/Manager

Analysis results relate only to the samples tested.

This document shall not be reproduced except in full, without the written approval of Liquilabs s.r.o.